

Field calibration for thermometers

TPI 341K FIELD CALIBRATION

Test the TPI advantage



Accessories

Adapters
Cables/Connectors
Oscilloscope Probes
Test Leads

Clamp-on meters (Current)

Combustion Efficiency Analyzers

Gas Detection

Carbon Monoxide (CO)
Combustibles
Refrigeration

Indoor Air Quality (IAQ)

Manometers (Pressure)

Multimeters (DMMs)

Oscilloscopes (Hand held)

Specialty Testers

Insulation
Multifunction
Photo-tachometer

Temperature

Contact
Non-Contact (IR)
Pocket Digital

Temperature Probes

J-Type
K-Type
T-Type
Thermistor

Test Leads & Accessories

Fused
Modular
Push-on
Screw-on

TPI

Headquarters:
9615 SW Allen Blvd.
Beaverton, OR 97005
USA
503-520-9197
Fax: 503-520-1225

The 341K has the ability to be field calibrated. Accuracy of $\pm 1^\circ\text{F}$ within the 30°F to 120°F range can be achieved by performing an ice bath calibration.

The temperature of an ice bath is approximately 32°F (0°C). Fill a plastic or metal container with crushed ice and add clean water to a depth of at least 4 inches. Stir the ice and water for 2 to 3 minutes prior to performing calibration to ensure the water is completely chilled. Make certain there is plenty of ice in the mixture and always use clean water. Distilled water works well.

1. Step-by-Step Procedures

1. Connect a temperature probe to the 341K.
2. Press and hold down the **MIN/MAX** and **HOLD** buttons and turn the 343 on. CALF will be displayed and the 343 will enter calibration mode and begin cycling between T1 and T2.
3. Insert the probes into the ice bath and allow the readings to stabilize.

Once the reading on both inputs has stabilized, press the **HOLD** button. SAVE will be displayed and calibration is complete.

